



Sensor	NTC
Accuracy	$\pm 0,5 \text{ }^{\circ}\text{C}$ (-20 $^{\circ}\text{C}$...+40 $^{\circ}\text{C}$) $\pm 0,8 \text{ }^{\circ}\text{C}$ for the remaining measurement range
Protection class	IP 65
Certificate	Factory calibration certificate
Measured variables	Temperature
Measuring range	-35 $^{\circ}\text{C}$...+75 $^{\circ}\text{C}$